

Calibration Report

Manufacturer: MEATROL ELECTRICAL

DATE	1 th Mar.2017
COIL MODEL	A-Y-FCT-200
QTY	3pcs

ORDER CODE	16PY02016
SERIAL NUMBER	
Rated Current	1000A

Secondary Side Reference Instrument

Model: KEITHLEY 2700
 Serial Number:
 Accuracy: ±0.50% of reading
 Calibrated by: MEATROL ELECTRICAL
 Calibration date: 01/10/14

Environmental Conditions during Calibration

Calibration performed in the temperature range:±18°C。。。+30°C
 Relative humidity not higher than : 60%

Standard output is 102mV(100% current)			
	INPUT CURRENT [A]	OUTPUT VOLTAGE [mV]	FREQUENCY [Hz]
REFERENCE INSTRUMENT READOUT	1000	101.8	60

Signature Jesse Jia
 2017. 3. 1

Approved Walter Long